

US-1B

Ultrasound Examination Training Model “ABDFAN”

Production supervision:
Junji Machi, MD, PhD
University of Hawaii at Manoa and Kuakini Medical Center

Caution:

Don't mark on the phantom with pen or leave printed materials contacted on its surface.
Ink marks on the phantom will be irremovable.

Instruction Manual

Contents



- Please read before training
 - General information P. 1
- Preparation
 - 1. Before training P. 2
- Training
 - 2. Training session P. 2
 - 3. After training P. 2

Unique high-fidelity ultrasound phantom facilitates effective training in abdominal ultrasound scanning with your own clinical devices.

Simulated lesions embedded as targets provide wider educational opportunities.

● Features

- Inanimate tool for training of a novice to demonstration by an expert.
Any ultrasound device with a convex probe can be used for the phantom.
- Detailed hepatobiliary, pancreatic and other abdominal anatomy meeting requirement for excellent training; Eight Couinaud's hepatic segments can be localized.
- Various simulated lesions including biliary stones and cysts, solid tumors (hypoechoic, hyperechoic and target-appearance) in the liver, pancreas, spleen and kidneys.
- Near real-size organs, structures and abnormal lesions.
- Durable long-life phantom materials.


Movie Site



English Site

Set includes

Before your first use, please ensure that you have all components listed below.



- a. Ultrasound abdominal phantom 1
the liver (segmental anatomy, portal and hepatic venous systems, ligamentum teres and ligamentum venosum), biliary tract (gallbladder, cystic duct, intrahepatic and extrahepatic bile ducts), pancreas (pancreatic duct), spleen, kidneys, detailed vascular structures (aorta, vena cava, celiac artery and its branches, portal vein and its branches, superior mesenteric vessels, renal vessels, etc)
- b. Positioning pillow 1
- c. DVD Tutorial 1
- d. talcum powder 1
- Instruction manual (this leaflet) 1

The setting of supine position and the lateral position is possible using a positioning pillow.



To maintain the body coat layer, lay down the phantom on the PILLOW included in the product, when scanning.
To enable easy maintenance and to enhance the durability, this phantom has a coat layer.
Friction between this coat layer and a board causes the damage to the coat layer, so **do not rotate the phantom on a hard board such as a desk.**



DO s

Handle with care
The materials for phantom and models are special composition of soft resin. Please handle with care at all times.

Cleaning and care
Please clean the phantom completely every time after you finish the training.
The remaining lubricating gel may deteriorate the phantom.

Keep the training set at room temperature, away from heat, moisture and direct sunlight.

DON'Ts

Never wipe the phantom or models with thinner or organic solvent.

Don't mark on the phantom with pen or leave printed materials contacted on their surface.
Ink marks on the models will be irremovable.

Please note: The color of the phantom may change over time, though, please be assured that this is not deterioration of the material and the ultrasonic features of the phantom stay unaffected.

This phantom has a thin layer of coating for protective purposes which may cause minor wrinkling. However, it has no effect on the ultrasound image.



Please watch the following movie about handling precautions of ultrasound phantoms.

<https://www.youtube.com/watch?v=D3VcaUYcPAY>

1 Before training



1. Set the phantom stably on the positioning pillows.



2. Spread lubricant gel on the phantom body.

2 Training session



The phantom can be scanned with actual ultrasound devices.

* The map of simulated targets is on the backside of this manual.

Examples of ultrasound images




Duplication lesion



Low echo target



Gallstone



High and non echo targets

3 After training



Wipe the remaining gel completely with wet cloth.
If remaining gel gets dried up on the body, it will be irremovable.




Caution

Don't mark on the phantom with pen or leave printed materials contacted on its surface.
Ink marks on the phantom will be irremovable.

Simulated lesions in liver, pancreas, kidney, spleen, gallbladder

- **Simulated stone** 3mm
- **Simulated stone** 5mm
- **Simulated cysts** 3mm
- **Simulated cysts** 10mm
- **Hypoechoic** 10mm
- **Simulated tumors** 3mm
- **Simulated tumors** 10mm
- **Simulated tumors with double edge** 20mm



The phantom includes:

the liver

(segmental anatomy, portal and hepatic venous systems, ligamentum teres and ligamentum venosum),

biliary tract

(gallbladder, cystic duct, intrahepatic and extrahepatic bile ducts),

pancreas

(pancreatic duct),

spleen,

kidneys,

detailed vascular structures

(aorta, vena cava, celiac artery and its branches, portal vein and its branches, superior mesenteric vessels, renal vessels, etc)

The contents of this manual are subject to change without prior notice.

No part of this instruction manual may be reproduced or transmitted in any form without permission from the manufacturer. Please contact manufacturer for extra copies of this manual which may contain important updates and revisions.

Please contact manufacturer with any discrepancies, typos, or mistakes in this manual or product feedback. Your cooperation is greatly appreciated.

 **KYOTO KAGAKU co.,LTD**

URL:<http://www.kyotokagaku.com>

e-mail:rw-kyoto@kyotokagaku.co.jp

■ Main Office and Factory

15 Kitanechoya-cho Fushimi-ku Kyoto 612-8388, Japan

Telephone : 81-75-605-2510

Facsimile : 81-75-605-2519

■ KyotoKagaku America Inc.

USA,Canada,and South America

3109 Lomita Boulevard, Torrance, CA 90505 ,USA

Telephone : 1-310-325-8860

Facsimile : 1-310-325-8867



15 Kitanechoya-cho Fushimi-ku

Kyoto, Japan 612-8388

Telephone : +81-75-605-2510

Fax : +81-75-605-2519

Dear customer,

Note

Thank you for choosing our ultra sound phantom.

There may be a case that the surface of phantom is in a wet condition.

The ultra sound image will not be affected. In such case, wipe the surface with medical use alcohol. **Do not wipe the surface with strong solvent such as benzene.**

If you have any questions regarding this matter, please do not hesitate to contact us at any time.

Faithfully yours,

Kyoto Kagaku Co., Ltd.

Email: rw-kyoto@kyotokagaku.co.jp